

Subject: While I have stopped the logging function by pressing the "Start/Stop" button, then, I cannot make the logger start recording data again by pressing again the "Start/Stop" button. Is this normal?

1. Refer to your question, yes, it's normal, the design is to prevent that user didn't download the full report before re-start logging.

2. For above reason, after user download the report, please double check and confirm the logger setting again by pressing "Save" & "OK" as attached image for your reference, then the logger can be activated by pressing "Start/Stop" button.

