



AZ Instrument Corp.

Web site: <http://www.az-instrument.com.tw>
E-mail: info@az-instrument.com.tw

Tel: 886-4-25326668 (rep.)
Fax: 886-4-25326593

Edit by Naima 30 Dec. 2019

Question

Is AZ's USB Datalogger compliant with FDA 21 CFR Part 11?

Reply:

Please note that the regulation of FDA 21 CFR Part 11 is for the software.

Due to AZ's USB datalogger don't required user to install any software, please note that it's not applicable, or you can say it's not complying with the FDA 21 CFR Part 11 regulation.