



AZ Instrument Corp.

Web site: <http://www.az-instrument.com.tw>
E-mail: info@az-instrument.com.tw

Tel: 886-4-25326668 (rep.)
Fax: 886-4-25326593

Edit by Naima 6 Feb. 2020

Question

Why the report shows all 70 degree C

Reply:

All 70 degree C means an error code, not the real measurement.

Please help to provide more detail information about the operating environment with test report for further problem shooting.