

Q: What is expiration date on the Certificate of Calibration of PDF Loggers?

Edit: Mari 08 Feb. 2021

Our customer is requesting an expiration date on the Certificate of Calibration for the PDF loggers, such as 88181 or 88160 series.

Answer:

We don't define the expiration date but only shows the calibration date. The expiration period is defined by customer's ISO system. For example, some companies define it as 2 years, some is 1 year and some more strict companies are 6 months. For this reason, we don't define the expiration date.