

Q: if its possible to use the Data Logger with its 4 AA internal batteries fitted at the same time the logger is powered with a 9V dc adapter via its dc input connector. The reason for this request is so I can replace the batteries with new ones without loosing any of the loggers setting up parameters due to loss of power on their removal.

Edit: Mari 2 Feb. 2022

Answer:

The recording can continue while you replace batteries but of course the adaptor must be connected during that period of time. Thank you very much.