



ETS DR. GENZ TAIWAN PS CO., LTD.
ACCREDITED TEST HOUSE

CERTIFICATE OF CONFORMITY

EU EMC - DIRECTIVE 2004/108/EC -

This certifies that the following designated product

pH/mV meter
Model No. : 8601
Brand name: AZ

.....
(Product identification)

complies with the essential protection requirements of Council Directive 2004/108/EC on the approximation of the laws of the Member States relating to electromagnetic compatibility.
This certificate is awarded following tests carried out on samples of the product referred to above.
Assessment of compliance of the product with the requirements relating to electromagnetic compatibility was based on the following standards :

EN 61326 : 2006,
CISPR11©IEC:2003+A1:2004,
IEC/EN 61000 - 4 - 2: 1995+A1: 1998+A2: 2001
IEC/EN 61000 - 4 - 3: 2002+A1: 2002

.....
(Identification of regulations / standards)

This certificate is issued for
AZ Instrument Corporation
6F-5 No.210, 38th Road Taichung Ind. Park, Taichung City, Taiwan R.O.C

.....
(Name / Address)

THE CERTIFICATION IS VALID ONLY IN ACCORDANCE WITH THE TEST REPORT NR. W6M20611-7573-E-11
THE CONFIRMATION IS VALID ONLY IF THE PRODUCT IS MANUFACTURED IN ACCORDANCE WITH THE TEST REPORT

GS LVD E/e-MARK FCC GSM UMTS CE-MARK SAR RADIO CB EMC DECT WLAN 3G Bluetooth R&TTE



This is the result of tests carried out on those samples of the product referred to above which were submitted for testing, in accordance with the specification for the respective standards. The certificate holder has the right to fix the ETS-MARK which shows the EMC segment onto each individual product in the series to which the inspection sample belongs, provided that the product is manufactured in accordance with the test report.



Dec 08, 2006

.....
(Date)


.....
K.C. Chen

ETS DR. GENZ TAIWAN PS CO., LTD.
6F, NO. 58, LANE 188, RUEY-KUANG RD., NEIHU, TAIPEI 114, TAIWAN, R.O.C.